



本製品の取り扱いについて

Part No. Z1-003-460, IA003591

Jul. 2005

USBインターフェースボード/GPIBインターフェースボード

US21 / IB21

本インターフェースボードは専用の製品に組み込んで使用しますので、本製品単体での取扱説明書は添付されておりません。本製品の組み込み方法、および操作方法については、本インターフェースボードが組み込まれる製品に添付されている取扱説明書をお読みください。

取り扱いの際には、以下の注意を必ず守ってください。

- 本製品に触れる前に、金属に触れて静電気を放電してください。
- 静電気の発生しやすい環境で本製品を取り扱わないでください。
- 保管するときは開封時の静電袋などに入れ、静電気対策を取ってください。
- 落下、衝撃を避けてください。
- 本インターフェースボードが組み込まれる製品の電源を入れたまま取り付け、取り外しをしないでください。

開梱時の点検

製品を受け取ったら、本製品が損傷していないかお確かめください。

万一、損傷がありましたら、購入先または当社営業所にお問い合わせください。

菊水電子工業株式会社

本社・技術センター 〒224-0023 横浜市都筑区東山田1-1-3

TEL: 045-593-0200(代)

<http://www.kikusui.co.jp>

Printed in Japan



HANDLING OF THE PRODUCT

Part No. Z1-003-460, IA003591

Jul. 2005

USB INTERFACE BOARD/ GPIB INTERFACE BOARD

US21 / IB21

The US21/IB21 interface board is installed in a dedicated product for operation. Therefore, an operation manual is not provided for this interface board by itself.

For the procedure of installing the interface board and the procedure for operating the interface board, read the attached manual to the product installed the interface board.

When handling the interface board, observe the following:

- Ground yourself by touching the chassis before touching the interface board.
- Avoid handling the interface board in an environment subject to strong static electricity.
- For storage, provide electrostatic protection measures such as the anti-static bag accompanying the interface board.
- Do not allow the interface board to be subjected to falls or impacts.
- Do not install or uninstall the interface board with the power ON of the product installed the interface board.

Checking the Package Contents

When you receive the US21/IB21, check that the interface board has not been damaged during transportation.

If the interface board is damaged, contact your Kikusui agent or distributor.

KIKUSUI ELECTRONICS CORP.

1-1-3, Higashiyamata, Tsuzuki-ku, Yokohama, 224-0023, Japan

Tel : +81-45-593-7570 Fax : +81-45-593-7571

<http://www.kikusui.co.jp>

Printed in Japan